

Declaration of Conformity

Responsible Manufacturer

Rexuma AB
 Västra Rydsvägen 122
 196 31 Kungsängen
 Sweden

Medical Device


Product	Article no.	Intended use
Vårdrumspanel	BHU	The Bed Head Unit is intended to be mounted, to be used in care facilities for general and intermediate intensive care in hospitals and other care facilities for supply of media service (electricity, gas, calling, data, illumination etc). The Bed Head Unit consists of horizontal and vertical aluminium profiles and different components/sockets (for electricity, gas, data etc). The units and the components can be combined in different ways according to the care facility's wishes about function from Rexuma's configuration list.

Conformity Assessment Procedure	Annex 5 Directive 93/42/EEC (LVFS 2003:11)
Notified Body	Rise NB 0402
EC Certificate No.	SC0099-17

Declaration of Conformity

I, the undersigned, hereby declare that Rexuma Vårdrumspanel is in conformity with the requirements of Directive 93/42/EEC and the Ministerial Decision of harmonization of national legislation, including the Swedish Medical Devices Act (1993:584) and LVFS 2003:11.

Signature



 Patrik Keussen, CEO

Place and date

Kungsängen 2022-12-14